



The FDA Safety Information and Adverse Event Reporting Program

FDA notified healthcare professionals and patients that certain transdermal patches (medicated patches applied to the skin), containing aluminum or other metals in the backing of the patches, can overheat during an MRI scan and cause skin burns in the immediate area of the patch. FDA is in the process of reviewing the labeling and composition of all medicated patches to ensure that those made with materials containing metal provide a warning about the risk of burns to patients who wear the patches during an MRI scan. Until this review is complete, FDA recommends that healthcare professionals referring patients to have an MRI scan identify those patients who are wearing a patch before the patients have the MRI scan. The healthcare professional should advise these patients about the procedures for removing and disposing of the patch before the MRI scan, and replacing the patch after the MRI scan. MRI facilities should follow published safe practice recommendations concerning patients who are wearing patches.

Read the MedWatch safety summary, including a link to the FDA Public Health Advisory, at:

<http://www.fda.gov/medwatch/safety/2009/safety09.htm#Transdermal>

You are encouraged to report all serious adverse events and product quality problems to FDA MedWatch at www.fda.gov/medwatch/report.htm

Update your subscriptions, modify your e-mail address, or stop subscriptions at any time on your [Subscriber Preferences Page](#). You will need to use your e-mail address to log in. If you have questions or problems with the subscription service, please contact support@govdelivery.com.

This service is provided to you at no charge by [U.S. Food & Drug Administration \(FDA\)](#).



GovDelivery, Inc. sending on behalf of U.S. Food & Drug Administration (FDA) · 5600 Fishers Lane · Rockville MD 20857 · 800-439-1420

GovDelivery, Inc. sending on behalf of U.S. Food & Drug Administration (FDA) · 5600 Fishers Lane · Rockville MD 20857 · 800-439-1420